



Declaration of product release Certification of conformity – C1

Manufacturer	Equflow BV - PO Box 35 - 5370 AA - The Netherlands
Product description	PVDF Disposable Rotor Tube 8.5; Hosebarb; Clipmounting
Part number	0085 PH000CX-CK001
Production Lot number	LOT 323
Lot release date	24-08-2017
Production date	03-10-2017
Serial number	0020666
Expire date *)	24-08-2020
Test engineer	SvB

Environmental test conditions:	
Medium	Demineralized water
Temperature	20° Celsius
Pressure	4 Bar (400 kPa)
K-factor **)	4625

Declares that above mentioned product:

Is/are manufactured, tested and calibrated in compliance with the manufacturer production and quality directives. The products are submitted by the manufacturer to the factory production control department that has performed the outgoing inspection.

This certificate attests that all provisions concerning the attestation of production control were applied.

Calibration equipment:	Make:	Model:	Serial #:	Cal. Expires:
Balance	BWT	BB-60-3000	72321014	16-06-2018
Calibrated by BWT. ISO17025 certificated				

Certificate number: 0020666
Accreditation by Equflow QA dept.

Date : 03-10-2017
Released by : Ronald de Lange
Function : Manager QA & Operations

Note:
To receive a copy of applicable Equflow Quality directives or part of it, description of the calibration or test procedure, or any other document in relation to validation or certification, contact your provider or send a request to sales@equflow.com to the attention of QA manager.
(*) Expire date is 36 months after the lot release date. If sterilization with gamma radiation is used, the expire date will be 24 months after gamma irradiation, but only if sterilization is conducted within 12 months of the lot release date and the maximum gamma dose of 50 kGy is not exceeded.
(**) K-factor: impulses per liter, based on water - established at 50% of flowrate at approximately 20°C
